

## UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF United States Patent and Trademark Address: COMMISSIONER FOR PATEN P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

FIRST NAMED INVENTOR ATTORNEY DOCKET NO. CONFIRMATION NO. APPLICATION NO. FILING DATE 12/28/2000 Clifton A. Alferness 1931-1 09/751,271 EXAMINER 7590 06/14/2004 ODLAND, KATHRYN P

GRAYBEAL, JACKSON, HALEY LLP 155 - 108TH AVENUE NE **SUITE 350** BELLEVUE, WA 98004-5901

PAPER NUMBER ART UNIT 3743

DATE MAILED: 06/14/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

## Supplemental Advisory Action

Application No.

O9/751,271

Examiner

Applicant(s)

ALFERNESS ET AL.

Art Unit

Kathryn Odland 3743

--The MAILING DATE of this communication appears on the cover sheet with the correspondence address -

THE REPLY FILED 22 April 2004 FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE. Therefore, further action by the applicant is required to avoid abandonment of this application. A proper reply to a final rejection under 37 CFR 1.113 may only be either: (1) a timely filed amendment which places the application in condition for allowance; (2) a timely filed Notice of Appeal (with appeal fee); or (3) a timely filed Request for Continued Examination (RCE) in compliance with 37 CFR 1.114.

condition for allowance; (2) a timely filed Notice of Appeal (with appeal fee); or (3) a timely filed Request for Continued Examination (RCE) in compliance with 37 CFR 1.114.
PERIOD FOR REPLY [check either a) or b)]
a) The period for reply expiresmonths from the mailing date of the final rejection.
b) The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection.  ONLY CHECK THIS BOX WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f).
Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).
1. A Notice of Appeal was filed on Appellant's Brief must be filed within the period set forth in 37 CFR 1.192(a), or any extension thereof (37 CFR 1.191(d)), to avoid dismissal of the appeal.
2. The proposed amendment(s) will not be entered because:
(a) They raise new issues that would require further consideration and/or search (see NOTE below);
(b) ☐ they raise the issue of new matter (see Note below);
(c) they are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal; and/or
(d) $\square$ they present additional claims without canceling a corresponding number of finally rejected claims.
NOTE:
3. Applicant's reply has overcome the following rejection(s):
4. Newly proposed or amended claim(s) would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s).
5. ☐ The a) ☐ affidavit, b) ☐ exhibit, or c) ☐ request for reconsideration has been considered but does NOT place the application in condition for allowance because: See Continuation Sheet.
6. The affidavit or exhibit will NOT be considered because it is not directed SOLELY to issues which were newly raised by the Examiner in the final rejection.
7. For purposes of Appeal, the proposed amendment(s) a) will not be entered or b) will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended.
The status of the claim(s) is (or will be) as follows:
Claim(s) allowed:
Claim(s) objected to:
Claim(s) rejected: 25-30.
Claim(s) withdrawn from consideration:
8. The drawing correction filed on is a) approved or b) disapproved by the Examiner.
9. Note the attached Information Disclosure Statement(s)( PTO-1449) Paper No(s)
10. Other:

Continuation of 5. does NOT place the application in condition for allowance because: The examiner has carefully reviewed applicant's arguments. Applicant argues, "The Houser device does not encircle the mitral valve at all while it is in the coronary sinus." However, the examiner respectfully disagrees. Applicant's attention is directed to figure 42B. Element 398 is the coronary sinus. The constrictive device such as 498a is advanced into the coronary sinus of the heart until the constrictive device at least partially encircles the mitral valve of the heart. In this figure, portion of the clip is encircling the valve and another portion is still within the coronary sinus. Applicant has failed to define over the prior art rejection. The claim language does not preclude penetrating through tissue.